

ISO 9001: 2015

SALT SPRAY TEST CHAMBER

This tester is designed for surface treatment of corrosion-resistant capacity test for various materials, including paint, electroplating, organic and inorganic film, anti-rust oil, and chemical treatment, and then achieves a long period of resistance to corrosion products.



FEATURES: -

- ➤ Automatic add water when water shortage, with a group of manual water
- > Special glass nozzle, guarantees non-crystalline block for 4000 hours
- With dual over-temperature protection, water shortage warning, ensure safety use
- > Temperature device adopts digital display, SSR control, precision is 0.1 centigrade degree
- > Chamber adopts direct steam heating method, fast warming, shortening standby time
- > Spray tower attached cone diffuser, with functions of directional mist, adjustable mist, and uniformity fog lamp.

Registered Office: UGF-12 Ansal Plaza-1 Chiranjeev Vihar Ghaziabad-201002 (U.P.) INDIA Workshop: Khasra No - 63 Village Bayana, Near Wave City, NH 24, Ghaziabad -201015 (U.P.) INDIA

Phone: +91- 82870 85201

Email: info@hsmleindia.in / sales@hsmleindia.com



ISO 9001: 2015

Technical Specifications

	1	1	1	
Model	HSC-108AT	HSC-270AT	HSC-480AT	HSC-800AT
Interior	600X450X400m	900X600X500mm	1200X850X500m	1600X1000X550
Dimensions (mm)	m		m	mm
Exterior dimensions	1130X630X1070	1460X910X1280	2200X1200X1450	2600X1450X1550
(mm)	mm	mm	mm	mm
Chamber Volume	108 L	270 L	480 L	800 L
Solution Tank	15L	25 L	40 L	40 L
Power Supply	AC 220V 50HZ			
Ambient temp. & humid. range	+5°C ~ +35°C / <85% RH			
Testing Temperature	NSS ACSS 35°C±1°C / CASS 50°C±1°C			
Air Temperature	NSS ACSS 47°C±1°C /CASS 63°C±1°C			
Solution Consistence	NaCl 5%			
PH Value	NASS ACSS:6.5~7.2 / CASS:3.0~3.2			
Temperature	≦±0.5°C			
fluctuation		4		
Temperature uniformity	≦2. 0°C			
Temperature deviation	≦±2.0°C			
Spray quantum	1.0~2.0(ml/80cm2/h)			
Chamber Material	polyvinyl chloride (PVC)			
Heater	Corrosion-resistant titanium tube			
Temperature sensor	Platinum Resistance (PT-100)			
Safety device	Over-temperature protection, leakage protection			

Registered Office: UGF-12 Ansal Plaza-1 Chiranjeev Vihar Ghaziabad-201002 (U.P.) INDIA Workshop: Khasra No - 63 Village Bayana, Near Wave City, NH 24, Ghaziabad -201015 (U.P.) INDIA

Phone: +91- 82870 85201

Email: info@hsmleindia.in / sales@hsmleindia.com



TESTING MACHINE STRUCTURE

MATERIALS USED: -

• Test chamber: use the PVC gray impact resistance plate (from Germany), 5mm in thickness;

ISO 9001 : 2015

- Thermostat water tank: use the gray impact resistance plate of PVC (from Germany), 5mm in thickness and it is equipped with the auto water level controller to avoid the inconvenience to refill the water by manpower.
- The upper cover of the test chamber: use the European transparent acrylic plate, 5mm in thickness;
- The barrel for refilling the reagent: equipped with a built-in water level gauge. Use the gray impact resistance
- plate made of PVC 5mm in thickness, which looks as an integral whole with the machine stand.
- The basket supporter inside the test chamber: use the gray PVC impact resistance plate.
- The test angle can be freely adjusted to 15° and 30°
 - 1) 48 test pieces can be placed by the angle of 15 degree
 - 2) 30 test pieces can be placed by the angle of 30 degree
- Note: the size of the test pieces is 150×70mm H×D
- The saturated air tank: made of stainless steel, 2mm in thickness and equipped with the auto water level @ controller to avoid the inconvenience to refill the water by manpower.

APPLICATION TEMPERATURE: -

- Set the type according to the standards of CNS, JIS and ASTM:
- The test method of salt water:
- o The test chamber: 35C ±1 C/
- o The saturated air barrel: 47C ±1 C ⊚
- The test method of corrosive resistant:
- o The test chamber: 50C ±1 C / The saturated air barrel: 63C ±1 C

Registered Office: UGF-12 Ansal Plaza-1 Chiranjeev Vihar Ghaziabad-201002 (U.P.) INDIA Workshop: Khasra No - 63 Village Bayana, Near Wave City, NH 24, Ghaziabad -201015 (U.P.) INDIA

Phone: +91- 82870 85201

Email: info@hsmleindia.in / sales@hsmleindia.com



ISO 9001: 2015

SPRAYING MODES: -

- Conical disperser: It can adjust the spraying and make the reek and falling reek even.
- Glass nozzles. The accurate glass nozzles (self-made)
- Note: volume of spraying: 1~2 ml / 80CM2 /1h (self-made)

CONTROL SYSTEM: -

- Temperature control of test chamber 0~99.9 with accuracy in 0.1 (LED with 2 figures) PID automatic calculator: 1 No.
- Safe temperature controller for liquid expansion in heating channel 0~110: 2
 Nos.
- The temperature controller of saturated air tank: 0~99.9 with accuracy in 0.1. (LED with 2 figures), able, PID automatic calculator: 1 No.
- Time controller: ANLY AH5R-2: 1 No.
- Total timer 0~9999.9 hr: 1 No.
- The first stage pressure adjusting structure:
- a) First pressure regulating valve with a set of pressure gauge (pre-adjust To 2~2.5kg/CM2)
- b) Second pressure regulating valve: pressure regulating valve 1/4" PT, pressure 0~4.0kg/ CM2, adjustable (adjust the spraying pressure to 1.0kg CM2).
- c) Pressure gauge 2kg/ CM2 (0.1kg/ CM2): 1 No.

SAFETY DEVICE

- · Water level controller: 1 Set
- State warning lamp: 3 Nos.
- Protection switch for over temperature: 2 Sets

STANDARD ACCESSORIES: -

- Bracket for 1 set (can carry the weight of test samples under 2kgs).
- 500g Nacl reagent contained in 2 bottles (contained in 4 bottles for standard type) © 10kg barrel for salt water making: 1 No.
- 80cm2 standard collecting cup: 1 No. (2pcs. for standard type) © 50ML graduated cylinder: 1 No. (2pcs. for standard type) © The detailed operational manual: 1 No.

Registered Office: UGF-12 Ansal Plaza-1 Chiranjeev Vihar Ghaziabad-201002 (U.P.) INDIA Workshop: Khasra No - 63 Village Bayana, Near Wave City, NH 24, Ghaziabad -201015 (U.P.) INDIA

Phone: +91- 82870 85201

Email: info@hsmleindia.in / sales@hsmleindia.com



FOR TOUCH SCREEN CONTROLLER FEATURES:

- AEE type 6.5-inch LCD touch screen control
- Digital circuit design, PID automatic calculation for temperature control
- Temp. resolution 0.1
- Touch modules can continuously use more than 200,000times
- Complete test function
- Dynamic display working interface, man-machine dialogue interface @ A buzzer alarm to signal the working condition of the machine @ Adopt ergonomic design, beautiful shape, easy operation.
- Power outage time memory function. Once the machine suddenly loses power or shut down, the test data
- and set data will be kept to the reboot.
- All function keys are integrated on the touch screen, all circuit adopts integrated circuit and modular @ structure, greatly reduce the incidence of failure.
- Lock screen function, operator can lock the screen when he need leave, it can prevent other persons to wrong operate the machine.



ISO 9001: 2015

Registered Office: UGF-12 Ansal Plaza-1 Chiranjeev Vihar Ghaziabad-201002 (U.P.) INDIA Workshop: Khasra No - 63 Village Bayana, Near Wave City, NH 24, Ghaziabad -201015 (U.P.) INDIA

Phone: +91-8750305291

Email: info@hsmleindia.in / sales@hsmleindia.com

Website: www.hopesmluckyenterprises.in / www.hsmleindia.com (Under Construction) website: www.nopesmluckyenterprises.in / www.nsmleindia.com (Under Construction)